

***VNA COMPARISON
AGILENT FIELDFOX N9923A (DEMO) AND AGILENT
E5061A***

***Leroy Johnson
Raul Monsalve***

SESE, Arizona State University

May 20, 2019

Description:

- There is a new VNA available in our lab, an AGILENT FIELDFOX N9923A, and it was compared to the AGILENT E5061A, we have been using until now.
- The comparison consisted of the accuracy of the measurement of four attenuators (3dB-BW-S3W2+, 6dB-BW-S6W2+, 10dB-BW-S10W2+, and 15dB-BW-S15W2+), after correction using the expectations for the open, short, and match (Male, 85033E). The temperature at all moments was $23.5 \pm 0.5^\circ\text{C}$
- With measurements specifications as follows:
 - i. Power: 0dBm
 - ii. bandwidth: 100 Hz (300 Hz Field-Fox: Lowest Setting)
 - iii. Max Frequency: 200MHz
 - iv. Min Frequency: 40MHz
 - v. AVG: 10 traces
 - vi. Points: 641
 - vii. Frequency Step (MHz): 0.250
- The steps in the testing at each VNA are:
 - i. Calibrate the VNA at its SMA port (with open, short, & match).
 - ii. Measure the S11 of the open, short, & match, AGAIN after calibration.
 - iii. Measure the S11 of the four attenuators.
 - iv. Repeat steps ii and iii for repeatability.
- Perform uncalibrated VNA measurements to determine uncorrected S11 parameters.
 - i. Measure open STD using uncorrected VNA measurement and save S1p_open
 - ii. Measure Short STD and save S1p_short
 - iii. Measure Load STD and save S1p_load
 - iv. Using the calibration coefficients of the OPEN, Short and load standards (found on the VNA, called C0, C1..., L0, L1...), we can compute the S parameters of a 2-port network called the "error network" (it is a hypothetical 2 port network between the VNA and DUT)
 - v. Measure the S11 of the DUT (attenuator) and "de-embed" the error network from the measured data, you will be left only with the S11 (complex) of the DUT.
 - vi. The equation you will need is this: (it is the S11 of a [S_error] network terminated with Gamma_dut)
$$S_{11_Measured} = S_{11_error} + (S_{21_Error} * \Gamma_{dut} * S_{12_error}) / (1 - S_{22_error} * \Gamma_{dut})$$
 - vii. Solve for the S parameters of the error network [S_error]: S11_error, S12_error, S21_error and S22_error by solving the above 4 unknowns equation for:
$$\Gamma_{dut} = \Gamma_{open_std}$$
$$\Gamma_{dut} = \Gamma_{short_std}$$
$$\Gamma_{dut} = \Gamma_{load_std}$$
 - viii. assume reciprocity: $S_{21_error} = S_{12_error}$ (reasonable assumption as the error network is reciprocal, meaning that $S_{12}=S_{21}$). The results will give 3 equations to solve for the 3 unknowns.

Images of VNA connections:




Figure1: AGILENT E5061A Calibration arrangement utilizing the male 85033E Calibration kit



Figure2: AGILENT FIELDFOX N9923A Calibration arrangement. utilizing the male 85033E Calibration kit

Results:




Figure3: Magnitude of the Open after calibration. AGILENT FIELDFOX N9923A measurement one brown trace, AGILENT FIELDFOX N9923A measurement two red trace, AGILENT_E5061A measurement one blue trace, and AGILENT_E5061A measurement two pink trace. The difference in repeatability between the two VNAs with the same settings is favoring the Agilent E5061A VNA. The AGILENT_E5061A have closer tolerances and consistent curvature.




Figure4: Magnitude of the Short after calibration. AGILENT FIELDFOX N9923A measurement one brown trace, AGILENT FIELDFOX N9923A measurement two red trace, AGILENT_E5061A measurement one blue trace, and AGILENT_E5061A measurement two pink trace. Both VNAs measurements are within .001dB from one another




Figure5: Magnitude of the Match after calibration. AGILENT FIELDFOX N9923A measurement one brown trace, AGILENT FIELDFOX N9923A measurement two red trace, AGILENT_E5061A measurement one blue trace, and AGILENT_E5061A measurement two pink trace. The difference in repeatability between the two VNAs with the same settings is favoring the AGILENT FIELDFOX N9923A VNA due to closer tolerances and less noise. However, the AGILENT FIELDFOX N9923A has better linear curvature.




Figure6: Magnitude of the 3dB Attenuator after calibration. AGILENT FIELDFOX N9923A measurement one brown trace, AGILENT FIELDFOX N9923A measurement two red trace, AGILENT_E5061A measurement one blue trace, and AGILENT_E5061A measurement two pink trace. The difference in repeatability between the two VNAs with the same settings is favoring the Agilent E5061A VNA due to closer tolerances and consistent curvature that is more predictable.




Figure7: Magnitude of the 6dB Attenuator after calibration. AGILENT FIELDFOX N9923A measurement one brown trace, AGILENT FIELDFOX N9923A measurement two red trace, AGILENT_E5061A measurement one blue trace, and AGILENT_E5061A measurement two pink trace. The difference in repeatability between the two VNAs with the same settings is favoring the Agilent E5061A VNA due to closer tolerances. Both VNAs have consistent curvature making calibrations more predictable.




Figure8: Magnitude of the 10dB Attenuator after calibration. AGILENT FIELDFOX N9923A measurement one brown trace, AGILENT FIELDFOX N9923A measurement two red trace, AGILENT_E5061A measurement one blue trace, and AGILENT_E5061A measurement two pink trace. The difference in repeatability between the two VNAs with the same settings is favoring the Agilent E5061A VNA due to closer tolerances. Both VNAs have consistent curvature making calibrations more predictable.




Figure9: Magnitude of the 15dB Attenuator after calibration. AGILENT FIELDFOX N9923A measurement one brown trace, AGILENT FIELDFOX N9923A measurement two red trace, AGILENT_E5061A measurement one blue trace, and AGILENT_E5061A measurement two pink trace. The difference in repeatability between the two VNAs with the same settings is favoring the AGILENT E5061A VNA due to closer tolerances. Both VNAs have consistent curvature making calibrations more predictable.

Discussion:

- The measurements with the two VNAs show good agreement
 - 3-dB attenuator (-5.584 dB), variance.010 dB
 - 6-dB attenuator (-11.54 dB), variance 0.005 dB
 - 10-dB attenuator (-19.69 dB), variance 0.3 dB
 - 15-dB attenuator (-30.1 dB), variance 0.7 dB
- Overall, most measurements are precise to one another varying slightly from one another.